

Certificate of Compliance

Certificate: 2379458

Master Contract: 234596

Project: 2379458

Date Issued: December 16, 2010

Issued to: EUROIMMUN Medizinische
Labordiagnostika AG
Seekamp 31
Lübeck, 23560
Germany

Attention: Dr. Winfried Stocker

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by: Celia Cordero
Celia Cordero

PRODUCTS

CLASS - 8721 05 LABORATORY EQUIPMENT Electrical

CLASS - 8721 85 ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

Flexible laboratory automation workstation, model Euroimmun Analyzer I-2P, rated 100 – 240VAC, 3.2 – 1.3A, 50 – 60Hz

Note: The above model is Cord connected, Equipment Class 1, Pollution Degree 2, Installation Category II.

APPLICABLE REQUIREMENTS

- | | | |
|-------------------------------------|---|--|
| CAN / CSA C22.2 No. 61010-1-04 | - | Safety requirements for Electrical Equipment for Measurement, Control and Laboratory use, Part1: General requirement |
| UL 61010-1, 2 nd Edition | - | Electrical Equipment for Laboratory use: Part1 General Requirements |

Certificate: 2379458

Project: 2379458



Master Contract: 234596

Date Issued: December 16, 2010

CONDITIONS OF ACCEPTABILITY

The equipment is supplied with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used.

MARKINGS

c CSA us Mark, Company's name or file number 235496, model designation, serial number, electrical rating – amperes, voltage, frequency, appear in a visible permanent manner.

The following markings appear on the product and may be described in further detail in the Certification Report Supplement, Section C:

- Company's name identification (company name and/or file number and/or registered tradename);
- Model designation;
- Electrical rating;
- Traceable serial number;
- The CSA Monogram with NRTL/C subscript;
- Replacement fuse markings;
- The IEC 417, No. 5019a earthing symbol;
- On and/or Off positions of the power switch;
- Convenience receptacle ratings;
- Laser markings;
- The ISO 7000 – 0434 caution symbol;
- The warning "DISCONNECT SUPPLY BEFORE SERVICING";



CSA INTERNATIONAL

Supplement to Certificate of Compliance

Certificate: 2379458

Master Contract: 234596

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
2379458	December 16, 2010	Original Certification. Submitter Model No. 233465 – 2042092 2 PS Listee Model No. Euroimmun Analyser I-2P

Certificate: 2379458

Project: 2379458



Master Contract: 234596

Date Issued: December 16, 2010
